

# Verification of Compliance

Product Name : Medical Type Switching Power Supply  
Model Number : NMP650-aaaa-xx(a=C,E,H,K,D or # ; x=0~9)  
C=NMS-240-05, E=NMS-240-12,  
H=NMS-240-24, K=NMS-240-48, D=NMD-240, #=Blank  
Applicant : MEAN WELL ENTERPRISES CO., LTD.  
Address : No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan  
(R.O.C.)  
Report Number : S2C-M020-1706-354  
Issue Date : December 22, 2020  
Applicable Standards :  
Emission : Immunity :  
EN 60601-1-2 :2015 EN 55024 : 2010+A1 :2015  
Home Healthcare Environment EN 60601-1-2 :2015 Home Healthcare Environment  
EN 55011 : 2016/A11:2020 EN 55035 : 2017+A11 :2020  
Group 1, Class B EN 61000-4-2 : 2009  
EN 55032: 2015/A11:2020 Class B EN 61000-4-3 : 2006+A1 : 2008+A2 : 2010  
EN IEC 61000-3-2 : 2019 EN 61000-4-4 : 2012  
EN 61000-3-3 : 2013+A1 :2019 EN 61000-4-5 : 2014+A1 :2017  
EN 61000-4-6 : 2014  
EN 61000-4-8 : 2010  
EN IEC 61000-4-11 : 2020



Based on the EMC Directive 2014/30/EU, Medical Device Directive 93/42/EEC (with it's modification) and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



## Central Research Technology Co.

11, Lane 41, Fushuen St., Jungshan Chiu,  
Taipei 104, Taiwan

Tel : 886-2-25984568

Fax : 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: December 22, 2020

TAF 0905

FCC CAB Code TW1104, TW0019

NVLAP Lab Code 200575-0

ISED CAB Code TW0905

VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010

T-20009, G-10614