



UK Declaration of Conformity

For the following equipment :

Product Name: Switching Power Supply

Model Designation: NMP650-aaaa-xx(a=C, D, E, H, K or #: x=0~9) C=NMS-240-05, E=NMS-240-12, H=NMS-240-24, K=NMS-240-48, #=Blank, D=NMD-240, #=Blank

The designated product(s) is(are) in conformity with the relevant legislation:

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012: SI 2012 No. 3032

Medical Devices Regulations 2002 (SI 2002 No 618, as amended)

Directive 93/42/EEC on medical devices (EU MDD)

Electrical Equipment (Safety) Regulations 2016 :

BS EN 62368-1:2014+A11:2017

BS EN 60601-1:2006+A12:2014

TUV certificate No : R50417981

TUV certificate No : TA50393537 (for NMP650)

TUV certificate No : TA50393494 (for NMS-240&NMD-240)

Electrical Compatibility Regulations 2016 :

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

BS EN 55032:2015+A11:2020

BS EN 55011:2016+A11:2020

Class B

Harmonic current BS EN IEC 61000-3-2:2019

Voltage flicker BS EN 61000-3-3:2013+A1:2019

EMS (Electro-Magnetic Susceptibility)

BS EN 55024:2010+A1:2015 BS EN 60601-1-2:2015 BS EN 55035:2017+A11:2020

ESD air BS EN 61000-4-2:2009 Level 4 15KV

ESD contact BS EN 61000-4-2:2009 Level 4 8KV

RF field susceptibility BS EN 61000-4-3: 2006+A1:2008+A2:2010 Level 3 10V/m (80MHz-2.7GHz)

RF field susceptibility BS EN 61000-4-3: 2006+A1:2008+A2:2010 Table 9 9-28V/m (385MHz-5.78GHz)

EFT bursts BS EN 61000-4-4: 2012 Level 3 2KV/5KHz

Surge susceptibility BS EN 61000-4-5:2014+A1:2017 Level 4 2KV/Line-Line

Surge susceptibility BS EN 61000-4-5:2014+A1:2017 Level 4 4KV/Line-Earth

Conducted susceptibility BS EN 61000-4-6:2014 Level 2 3V

Magnetic field immunity BS EN 61000-4-8:2010 Level 4 30A/m

BS EN IEC 61000-4-11:2020

Voltage dip, interruption >95% dip 0.5 periods 30% dip 25 periods >95% interruptions 250 periods

Note:

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Regulations on the complete system again.

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

This Declaration is effective from serial number TC1xxxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

Taiwan

(Place)

Aries

(Signature)

May. 27th, 2021

(Date)

Alex Tsai/ Director, Product Strategy Center :

(Name / Position)

(Signature)