

Verification of Compliance

Product Name : AC/DC SWITCHING ADAPTOR
Brand Name : MEAN WELL
Model Number : GSM18xy (x=B,E; y=05, 07, 09,12, 15, 18, 24, 48),
GSM25xy (x=B,E; y=05, 07, 09,12, 15, 18, 24, 48),
GSM36xy (x=B,E; y=05, 07, 09,12, 15, 18, 24, 48)
Applicant : MEAN WELL ENTERPRISES CO., LTD.
Address : No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,
Taiwan (R.O.C.)
Report Number : C-M020-1304-409
Issue Date : December 2, 2013
Applicable Standards :

Emission:

EN 55011: 2009+A1:2010
Group 1, Class B
EN 55022: 2010 Class B
EN 61000-6-3: 2007+A1:2011
EN 61000-3-2: 2006+A1:2009+A2:2009
EN 61000-3-3: 2008
EN 61204-3: 2000 Class B

Immunity:

EN 55024: 2010
EN 61000-4-2: 2009
EN 61000-4-3: 2006+A1: 2008+A2: 2010
EN 61000-4-4: 2012
EN 61000-4-5: 2006
EN 61000-4-6: 2009
EN 61000-4-8: 2010
EN 61000-4-11: 2004

EN 61204-3: 2000: Low severity level
EN 61000-4-2: 2009
EN 61000-4-3: 2006+A1:2008+A2:2010
EN 61000-4-4: 2012
EN 61000-4-5: 2006
EN 61000-4-6: 2009
EN 61000-4-11: 2004

EN 61000-6-1: 2007
EN 61000-4-2: 2009
EN 61000-4-3: 2006+A1:2008+A2:2010
EN 61000-4-4: 2012
EN 61000-4-5: 2006
EN 61000-4-6: 2009
EN 61000-4-8: 2010
EN 61000-4-11: 2004

EN 60601-1-2: 2007 Life-supporting Equipme
EN 61000-4-2: 2009
EN 61000-4-3: 2006+A1:2008+A2:2010
EN 61000-4-4: 2012
EN 61000-4-5: 2006
EN 61000-4-6: 2009
EN 61000-4-8: 2010
EN 61000-4-11: 2004

Based on the EMC Directive 2004/108/EC, Medical Device Directive 93/43/EEC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905

FCC CAB Code TW1053

NVLAP Lab Code 200575-0

IC Code 4699A

VCCI Accep. No. R-1527, C-1609, T-1441, G-10,
C-4400, T-1334, G-614



Central Research Technology Co.

EMC Test Laboratory

11, Lane41, Fushuen St., Jungshan Chlu,
Taipei, Taiwan, 104, R.O.C.

Tel: 886-2-25984568

Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: December 2, 2013