

Certificate



Group Asia

Certificate no.

TA 50082469 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Eljintek, Inc.
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Test report no.: ZTW1-BHU 10016492 001

Client Reference: 0603028002/LDI

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2


Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation : MES50X-Y(MW)
X stands for : A or C
Y stands for : 0, 1, 2, 3, 4, 5, 6, 7, 8 or 1-1
Rated Voltage : AC 100-240V, 50/60Hz
Rated Current : 1.5-0.8A
Outputs : see constructional dataform
max. Ambient Temperature: 40°C
Protection Class : I

8
1
1


Remark: Primary and secondary circuits are separated according to method 4 of clause 17.g.



10

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue

(day/mo/yr)

07/04/2006

Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker