

Certificate



Certificate no.

TA 50388097 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 50080048 001

Client Reference: FL-21042017-TUV/MFM/

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

Type Designation	: MFM-30-x (MW)	8
x stands for	: 3.3, 5, 12, 15, 24, 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.75-0.5A	
Outputs	: see constructional dataform	
max. Ambient Temperature	: 60°C	
max. Operational Altitude	: 5000m	
Protection Class	: II	

Remarks: 1) The risk management requirements and usability
of the standard were not addressed;
2) This certificate is valid for maximum 5 years.

9

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419050807

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

07/09/2017

Certificate



Certificate no.

TA 50388097 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 50080048 001

Client Reference: FL-21042017-TUV/MFM/

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: MPM-30-xy (MW)	1
x stands for	: 3.3, 5, 12, 15, 24, 48	1
y stands for	: ST or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.75-0.5A	
max. Operational Altitude:	3000m (MPM-30-xST), 5000m (MPM-30-x)	

This certificate is valid for maximum 5 years.

Appendix: 1

3

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419050807

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

07/09/2017

Certificate



Certificate no.

TA 50388097 03

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P. R. China

Test report no.: ZTW1- 50080048 001

Client Reference: FL-21042017-TUV/MFM/

Tested to:

IEC 60601-1:2005+A1

EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:




IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419050807

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

07/09/2017

Certificate



Certificate no.

TA 50388097 04

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town,
Huadu District, Guangzhou 510890
P. R. China

Test report no.: ZTW1- 50080048 001

Client Reference: FL-21042017-TUV/MFM/

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419050807

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

**Date of Issue
(day/mo/yr)**

07/09/2017